

# ARCS Instrument/Component Replacement Form

Fill out the following form when replacing instruments at an ARCS site: Darwin – ARCS-3

SV-

System: Skyrad

Site: Darwin

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) $\times 10^{-6}$	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
Brusag Tracker	# 16 WD24216	Skyrad	n/a	n/a	n/a	27 May 06	01:30	05:35

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) $\times 10^{-6}$	Offset	Copy Config. Logger to PC
SciTech Tracker	00189	Skyrad	n/a	n/a	n/a	n/a

Replacement completed by: John Glowacki

## **NOTES:**

Old Part remains on site as spare: Yes

Old part packed for shipment to:

Other old part disposition (explain): Serviceable

Other comments: **Brusag Tracker was wandering for a brief period around solar noon (possibly a set-up problem).**